

**The Appendix is an integral part of
Certificate of Accreditation No. 64/2023 of 09/02/2023**

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

TRIGON PLUS s.r.o.
CAB number 2377, TRIGON PLUS, Calibration Laboratory
Západní 93, 251 01 Čestlice

Field of measured quantity: Temperature

Ordinal number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the measurand	Lowest stated expanded measurement uncertainty ²	Calibration principle	Calibration procedure identification ³	Work place
		min	unit	max	unit					
1*	Indication thermometers including temperature measuring chains	-80 °C	up to	-60 °C			0.49 °C	Comparison with a reference digital thermometer in a freezer box	KL-PP-01	
		-60 °C	up to	-20 °C			0.27 °C			
		-20 °C	up to	60 °C			0.13 °C			

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M a part of CMC and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95 %. If not stated otherwise, the uncertainty values stated without a unit are relative to the measured value. The uncertainty value stated herein is based on the best conditions achievable by the laboratory; the uncertainty value of a specific calibration may be higher depending on the conditions of such a calibration. For identical extreme values of adjacent ranges, the lower uncertainty value always applies.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).